

1 DMID Note-revised version

2

3 **CHROMagar™ ESBL / mSuperCARBA bi-plate medium**
4 **for detection of ESBL- and carbapenemase-producing**
5 ***Enterobacteriaceae from spiked stools***

6

7 **Delphine GIRLICH^a, Victor GROSPERRIN^a, Thierry NAAS^{a,b} and Laurent**
8 **DORTET^{a,b*}**

9

10 ^a EA7361, Université Paris-Sud, Bacteriology-Hygiene unit, APHP, Hôpital
11 Bicêtre, Le Kremlin-Bicêtre, France

12 ^b Associate French National Reference Center for Antibiotic Resistance
13 “Carbapenemase-producing Enterobacteriaceae”

14

15 *Corresponding author: Dr. Laurent DORTET

16 Faculté de Médecine et Université Paris Saclay, Le Kremlin-Bicêtre, 78 rue du
17 Général Leclerc 94270 Le Kremlin-Bicêtre, France.

18 Phone :+33 1 45212019

19 Fax : +33 1 45216340

20 Email: laurent.dortet@aphp.fr

21

22 **Running title :** Bi-plate for screening ESBL-Es and CPEs

23 **Word count :** Text : 1282 words; Abstract : 49 words

24

25

ABSTRACT

26 The recently released CHROMagar™ ESBL/CHROMagar™ mSuperCARBA bi-
27 plate medium was evaluated for the detection of ESBL- and carbapenemase-
28 producing *Enterobacteriaceae*. Spiked stools were used to mimic *in vivo* stool
29 colonization. Two-hundred enterobacterial isolates were tested. Respective
30 sensitivities of 93.9% and 97.8% were obtained for the detection of ESBL and
31 carbapenemase producers.

32

33 **Keywords :** carbapenemases, NDM, sensitivity, specificity, selective medium,
34 chromogenic

35

36

37 Resistance to broad-spectrum cephalosporins and subsequently to
38 carbapenems spreads rapidly among *Enterobacteriaceae* at an alarming rate,
39 resulting in both nosocomial and community-acquired infections primarily due to
40 the production of extended-spectrum β -lactamases (ESBLs) and/or
41 carbapenemases (1, 2). Targeted surveillance of high-risk patients and
42 screening are essential to prevent outbreaks of nosocomial infections by these
43 organisms. In *Enterobacteriaceae*, clinically relevant carbapenemases belong to
44 the Ambler class A (KPC-type), Ambler class B metallo- β -lactamases (NDM,
45 VIM and IMP), and Ambler class D enzymes (OXA-48-like) (3). The level of
46 resistance to carbapenems conferred by these carbapenemase producers may
47 vary widely, making their detection difficult when just based on their *in vitro*
48 susceptibility profile (4). ESBL are enzymes that mediate resistance to
49 penicillins, extended-spectrum third generation cephalosporins (3GC) and
50 monobactams. ESBL-producing *Enterobacteriaceae* appeared in the 1980s,
51 and have since become highly prevalent in nosocomial infections, mainly due to
52 *Escherichia coli* and *Klebsiella* spp. ChromID ESBL (bioMérieux) and
53 CHROMagar™ ESBL (CHROMagar, Paris, France) are very useful for the
54 detection of ESBL enterobacterial producers, nevertheless their ability to detect
55 isolates producing OXA-48-like carbapenemase alone is limited (5, 6).
56 Consequently, the detection of all carbapenemase-producing
57 *Enterobacteriaceae* (CPE), including OXA-48-like producers, is based on the
58 inoculation of at least two media able to efficiently identify (i) all types of
59 carbapenemases with the exception of OXA-48-like enzymes (*e.g.* ChromID®
60 CARBA medium, bioMérieux) and (ii) specifically OXA-48-like producers (*e.g.*
61 ChromID® OXA-48 medium, bioMérieux). These two media are now marketed
62 in a single bi-plate, the ChromID® CARBA SMART medium that has been

63 reported to efficiently detect all CPEs. However, this medium is designed for the
64 screening of CPEs and not for ESBL-producers.

65 Another medium had already been developed for the detection of all CPEs,
66 including OXA-48-like carbapenemase-producers, the SUPERCARBA medium
67 (7). The initial SUPERCARBA medium was further modified to improve its
68 stability and facilitate the presumptive identification of the enterobacterial
69 species by the addition of chromogenic molecules. This improved medium,
70 named mSuperCARBA (CHROMagar Ltd., France), was demonstrated to be
71 reliable for the screening of CPE (8-11). However, as observed with the
72 ChromID® CARBA SMART medium, this single plate medium is not suitable for
73 screening of ESBL producers.

74 Recently, the selective bi-plate medium CHROMagar™ ESBL/CHROMagar™
75 mSuperCARBA (CHROMagar Ltd., France) was designed for the rapid
76 identification of ESBL-producing and carbapenem-resistant *Enterobacteriaceae*.
77 The aim of the present study was to evaluate the performance of this bi-plate
78 medium using spiked stool samples with a collection of well-characterized CPE
79 and ESBL-producers at the molecular level.

80

81 Two hundred and three enterobacterial isolates were tested (Table 1),
82 including 137 carbapenemase-producers [30 Ambler class A (KPC, Sme, IMI,
83 GES-5, FRI-1), 50 Ambler class B (IMP, VIM, GIM-1, NDM-type), 43 Ambler
84 class D (OXA-48-type and OXA-372), 14 multiple carbapenemase producers]
85 and 62 non carbapenemase-producers [ESBL with or without impaired outer-
86 membrane (n=26), cephalosporinase-producers with or without an altered
87 outer-membrane (n=27), concomitant ESBL and over-expressed
88 cephalosporinase-producers (n=4), K1 penicillinase or extended-spectrum
89 oxacillinases-producers (n=5)]. The β-lactamase content of all these isolates

90 was characterized by-end-point PCR followed by Sanger sequencing or by
91 whole genome sequencing. In addition the Carba NP test was also performed
92 on all isolates as previously described (12). Fully-susceptible strains of *E. coli*
93 ATCC 25922 and *K. pneumoniae* CIP53153, and ESBL-producing strains of *E.*
94 *coli* CIP103983 (TEM-4) and *K. pneumoniae* ATCC 700603 (SHV-18) were
95 used as controls (not included in the Table of results).

96 CHROMagar™ ESBL/mSuperCARBA is a ready-to-use bi-plate with a
97 shelf life of 60 days at +4°C (+2°C to +12°C). Bacterial suspensions with a 0.5
98 McFarland (inoculum of ~5 x 10⁷ to 10⁸ CFU/mL) of the different isolates were
99 serially diluted (10-fold dilutions) in water. Spiked fecal samples were prepared
100 by adding 100 µL of each dilution to 900 µL of fecal suspension obtained by
101 resuspending 4 g of freshly pooled feces from four healthy volunteers in 40 mL
102 of distilled water, as previously described (13). A fecal suspension without
103 spiking was used as negative control. The lowest detection limits were
104 determined by plating 10 µL of each spiked fecal samples on CHROMagar™
105 ESBL/mSuperCARBA medium. Viable bacteria were counted after 24 h of
106 culture at 37°C. The lowest detection limit (LOD) corresponds the minimum
107 number of bacteria that must be present in the sample to obtain a growth on the
108 selective medium. In order to compare both selective media, the cut-off value
109 was arbitrarily set at > 1 x 10⁴ CFU/mL. A minimal concentration greater than 1
110 x 10⁴ CFU/mL was considered as non-detection and sensitivity and specificity
111 values were determined accordingly (Table 1).

112 Considering the 62 non-carbapenemase producers, all ESBL-producers
113 with (n=13) or without an altered outer-membrane (n=13) grew on the
114 CHROMagar™ ESBL medium, while 75% of the enterobacterial isolates with an
115 acquired cephalosporinase did not grow with the exception of one DHA-2-
116 producing *K. pneumoniae* isolate (considered as a false positive). However,

117 19/23 isolates with overexpression of the cephalosporinase associated with
118 decreased altered outer membrane grew on the CHROMagar™ ESBL medium.
119 They were falsely considered as potential ESBL-producers. Only one of the
120 three extended-spectrum oxacillinase-producers grew on the CHROMagar™
121 ESBL medium (thus giving two false negatives). Finally, the two penicillinase K1
122 overexpressing *K. oxytoca* were also falsely detected as ESBL-producers.
123 Overall, the sensitivity of the CHROMagar™ ESBL medium for the detection of
124 true ESBL-producers was 93.9% [95% confidence interval (CI95) = 78.4% -
125 98.9%] in this study. Due to the high number of isolates with overexpressed-
126 cephalosporinase associated with decreased outer-membrane permeability,
127 and the absence of isolates with wild-type phenotypes, the specificity of the
128 CHROMagar™ ESBL medium was low (24.1%). Despite this limitation in the
129 number of susceptible isolates tested, several studies have previously
130 demonstrated that this CHROMagar™ ESBL medium is useful for the efficient
131 detection of ESBL producers (14, 15).

132 Among the 137 CPEs, all but three cultured on the CHROMagar™
133 mSuperCARBA medium. One Sme-1-producing *Serratia marcescens*, one
134 NDM-1 producing *Providencia stuartii*, and one VIM-1 producing *E. coli* (Table
135 1) did not grow on the ‘carba side of the plate’, however, NDM-1 producing *P.*
136 *stuartii* and VIM-1 producing *E. coli* grew on the ESBL side of the plate. As
137 previously described with the SUPERCARBA medium, all OXA-48-like
138 producers, known to have a lower carbapenemase activity compared to other
139 carbapenemases (KPC, NDM, VIM, IMP), were detected irrespective of the
140 production of an associated ESBL (7). Despite high MICs of carbapenems,
141 most (15/23) non carbapenemase-producers with overexpression of the
142 cephalosporinase associated with decreased outer membrane permeability did
143 not culture on mSuperCARBA medium (Table 1). Overall, the sensitivity of the

144 CHROMagar™ mSuperCARBA medium for the detection carbapenemase
145 producers was 97.8% [CI95 = 93.2% - 99.4%] in this study. The specificity was
146 66.1% [CI95 = 52.9% - 77.4%], which corresponds mainly to the growth of
147 isolates with reduced susceptibility to carbapenems due to ESBL production
148 and/or AmpC overexpression plus impermeability of the outer membrane (Table
149 1). Due to the low specificity of both sides of the bi-plate, confirmatory tests for
150 ESBLs or carbapenemases detection should be undertaken on any growth. This
151 is a major drawback that will increase the turn around time and costs if a rapid
152 test is implemented as a confirmation test. However, this low specificity is mostly
153 due to the design of the strains collection and might be better from clinical
154 samples.

155

156 Our results suggest that the CHROMagar™ ESBL/mSuperCARBA bi-
157 plate might be a reliable tool for the concomitant screening of patients colonized
158 with ESBL and/or CPE. Currently, in France, hospitalized patients who have
159 travelled abroad are screened for the presence of CPE and ESBL-producing
160 Enterobacteriaceae. Different infection-control measures are implemented and
161 maintained according to the screening results (16): (i) standard precautions [no
162 ESBL producer, no CPE and no glycopeptide resistant *Enterococcus faecium*
163 (GRE)], (ii) contact precautions without cohorting (presence of ESBL producer
164 only), and (iii) ‘extensively drug-resistant bacteria’ (eXDR) control program that
165 includes dedicated staff, isolation of carrier and even cohorting, if possible, and,
166 active contact tracing and screening. This active screening of eXDR involves
167 the inoculation of three to four different media including one for the detection of
168 ESBL producers, one (bi-plate) or two media for the detection of CPE and one
169 for the screening of GRE. In the field of laboratory automation in clinical
170 microbiology (17, 18), the panel of media that can be loaded onto the sorter of

171 the inoculation system is limited (19). Accordingly, the implementation of
172 CHROMAGART™ ESBL/mSuperCARBA bi-plate might be an advantage, since
173 this medium can replace two or three of the currently used media for the
174 screening of ESBL and CPE producers. The cost of these bi-plates will be
175 approximately 3.25€/plate.

176

177 **ACKNOWLEDGMENTS**

178 We thank Saoussen Oueslati and Sandrine Bernabeu for helping constituting the
179 panel of isolates.

180 **FUNDING**

181 This work was partially supported by the University Paris-Sud, France and
182 Assistance Publique - Hôpitaux de Paris, Paris, France. LD, DG and TN are
183 members of the Laboratory of Excellence in Research on Medication and
184 Innovative Therapeutics (LERMIT) supported by a grant from the French National
185 Research Agency (ANR-10-LABX-33).

186

187 **TRANSPARENCY DECLARATION**

188 None to declare

189

REFERENCES

- 191 1. **Pitout JD, Laupland KB.** 2008. Extended-spectrum beta-lactamase-
192 producing Enterobacteriaceae: an emerging public-health concern.
193 Lancet Infect Dis **8**:159-166.
- 194 2. **Rood IGH, Li Q.** 2017. Review: Molecular detection of extended
195 spectrum-beta-lactamase- and carbapenemase-producing
196 Enterobacteriaceae in a clinical setting. Diagn Microbiol Infect Dis
197 **89**:245-250.
- 198 3. **Nordmann P, Dortet L, Poirel L.** 2012. Carbapenem resistance in
199 Enterobacteriaceae: here is the storm! Trends Mol Med **18**:263-272.
- 200 4. **Landman D, Urban C, Backer M, Kelly P, Shah N, Babu E, Bratu S,**
201 **Quale J.** 2010. Susceptibility profiles, molecular epidemiology, and
202 detection of KPC-producing Escherichia coli isolates from the New York
203 City vicinity. J Clin Microbiol **48**:4604-4607.
- 204 5. **Carrer A, Fortineau N, Nordmann P.** 2010. Use of ChromID extended-
205 spectrum beta-lactamase medium for detecting carbapenemase-
206 producing Enterobacteriaceae. J Clin Microbiol **48**:1913-1914.
- 207 6. **Hornsey M, Phee L, Woodford N, Turton J, Meunier D, Thomas C,**
208 **Wareham DW.** 2013. Evaluation of three selective chromogenic media,
209 CHROMagar ESBL, CHROMagar CTX-M and CHROMagar KPC, for the
210 detection of Klebsiella pneumoniae producing OXA-48 carbapenemase.
211 J Clin Pathol **66**:348-350.
- 212 7. **Nordmann P, Girlich D, Poirel L.** 2012. Detection of carbapenemase
213 producers in Enterobacteriaceae by use of a novel screening medium. J
214 Clin Microbiol **50**:2761-2766.
- 215 8. **Garcia-Fernandez S, Hernandez-Garcia M, Valverde A, Ruiz-
216 Garbajosa P, Morosini MI, Canton R.** 2017. CHROMagar

- 217 mSuperCARBA performance in carbapenem-resistant
218 Enterobacteriaceae isolates characterized at molecular level and routine
219 surveillance rectal swab specimens. Diagn Microbiol Infect Dis **87**:207-
220 209.
- 221 9. **Garcia-Quintanilla M, Poirel L, Nordmann P.** 2018. CHROMagar
222 mSuperCARBA and RAPIDEC(R) Carba NP test for detection of
223 carbapenemase-producing Enterobacteriaceae. Diagn Microbiol Infect
224 Dis **90**:77-80.
- 225 10. **Genc O, Aksu E.** 2018. Chromogenic culture media or rapid
226 immunochromatographic test: Which is better for detecting Klebsiella
227 pneumoniae that produce OXA-48 and can they be used in blood and
228 urine specimens. J Microbiol Methods **148**:169-173.
- 229 11. **Soria Segarra C, Larrea Vera G, Berrezueta Jara M, Arevalo Mendez
230 M, Cujilema P, Serrano Lino M, Soria Segarra C.** 2018. Utility of
231 CHROMagar mSuperCARBA for surveillance cultures of
232 carbapenemase-producing Enterobacteriaceae. New Microbes New
233 Infect **26**:42-48.
- 234 12. **Dortet L, Agathine A, Naas T, Cuzon G, Poirel L, Nordmann P.** 2015.
235 Evaluation of the RAPIDEC(R) CARBA NP, the Rapid CARB Screen(R)
236 and the Carba NP test for biochemical detection of carbapenemase-
237 producing Enterobacteriaceae. J Antimicrob Chemother **70**:3014-3022.
- 238 13. **Naas T, Ergani A, Carrer A, Nordmann P.** 2011. Real-time PCR for
239 detection of NDM-1 carbapenemase genes from spiked stool samples.
240 Antimicrob Agents Chemother **55**:4038-4043.
- 241 14. **Grohs P, Tillecovidin B, Caumont-Prim A, Carbonnelle E, Day N,
242 Podglajen I, Gutmann L.** 2013. Comparison of five media for detection

- 243 of extended-spectrum Beta-lactamase by use of the wasp instrument for
244 automated specimen processing. *J Clin Microbiol* **51**:2713-2716.
- 245 15. **Saito R, Koyano S, Nagai R, Okamura N, Moriya K, Koike K.** 2010.
246 Evaluation of a chromogenic agar medium for the detection of extended-
247 spectrum ss-lactamase-producing Enterobacteriaceae. *Lett Appl
248 Microbiol* **51**:704-706.
- 249 16. **Fournier S, Desenfant L, Monteil C, Nion-Huang M, Richard C,
250 Jarlier V, The Ap-Hp Outbreaks Control G.** 2018. Efficiency of different
251 control measures for preventing carbapenemase-producing
252 enterobacteria and glycopeptide-resistant *Enterococcus faecium*
253 outbreaks: a 6-year prospective study in a French multihospital
254 institution, January 2010 to December 2015. *Euro Surveill* **23**.
- 255 17. **Croxatto A, Greub G.** 2017. Project management: importance for
256 diagnostic laboratories. *Clin Microbiol Infect* **23**:434-440.
- 257 18. **Lina G, Greub G.** 2016. Automation in bacteriology: a changing way to
258 perform clinical diagnosis in infectious diseases. *Clin Microbiol Infect*
259 **22**:215-216.
- 260 19. **Croxatto A, Prod'hom G, Faverjon F, Rochais Y, Greub G.** 2016.
261 Laboratory automation in clinical bacteriology: what system to choose?
262 *Clin Microbiol Infect* **22**:217-235.
- 263

Table 1. Sensitivity of detection of the CHROMagar™ ESBL /mSuperCARBA medium from spiked fecal samples.

β-lactam resistance mechanism	Name of the strain	Species	MIC (µg/ml)			β-lactamase content	Limit of detection (UFC/ml) ^{b,c}	
			IPM ^a	ETP	MEM		ESBL	mSuperCARBA
NON CARBAPENEMASE PRODUCERS								
Acquired Case	2 E2	<i>E. coli</i>	0.12	0.02	0.02	DHA-1	> 1 x 10 ⁴	> 1 x 10 ⁴
	2 E3	<i>E. coli</i>	0.12	0.12	0.12	ACC-1	> 1 x 10 ⁴	> 1 x 10 ⁴
	2 E4	<i>K. pneumoniae</i>	0.12	0.5	0.12	DHA-2	1 x 10 ²	> 1 x 10 ⁴
	2 E5	<i>P. mirabilis</i>	0.25	0.12	0.12	ACC-1	> 1 x 10 ⁴	> 1 x 10 ⁴
ESBL	2 E6	<i>E. coli</i>	0.12	0.12	0.12	CTX-M-1	1 x 10 ²	> 1 x 10 ⁴
	2 E7	<i>E. coli</i>	0.12	0.12	0.12	CTX-M-3	1 x 10 ²	> 1 x 10 ⁴
	2 E8	<i>K. pneumoniae</i>	0.12	0.12	0.12	CTX-M-3	1 x 10 ²	> 1 x 10 ⁴
	2 E9	<i>E. coli</i>	0.12	0.12	0.12	CTX-M-14	1 x 10 ²	> 1 x 10 ⁴
	2 E10	<i>E. coli</i>	0.12	0.12	0.12	CTX-M-14	1 x 10 ²	> 1 x 10 ⁴
	2 F1	<i>K. pneumoniae</i>	0.12	0.12	0.12	CTX-M-14	1 x 10 ²	> 1 x 10 ⁴
	2 F2	<i>E. coli</i>	0.12	0.12	0.12	CTX-M-15	1 x 10 ²	> 1 x 10 ⁴
	2 F3	<i>E. coli</i>	0.12	0.12	0.12	CTX-M-15	1 x 10 ²	> 1 x 10 ⁴
	2 F4	<i>K. pneumoniae</i>	0.12	0.12	0.12	CTX-M-15	1 x 10 ²	> 1 x 10 ⁴
	2 F5	<i>K. pneumoniae</i>	0.12	0.12	0.12	CTX-M-15	1 x 10 ²	> 1 x 10 ⁴
Case + impermeability	2 F6	<i>K. pneumoniae</i>	0.12	0.12	0.12	CTX-M-15	1 x 10 ²	> 1 x 10 ⁴
	2 F7	<i>E. cloacae</i>	0.12	0.12	0.12	CTX-M-15	1 x 10 ²	> 1 x 10 ⁴
	2 F8	<i>E. cloacae</i>	0.12	0.12	0.12	VEB-1	1 x 10 ²	> 1 x 10 ⁴
	2 F9	<i>E. coli</i>	16	>32	2	■■■ Case	3 x 10 ²	2 x 10 ²
	CNR 92 10	<i>E. cloacae</i>	0.19	1	0.12	■■■ Case	1 x 10 ²	> 1 x 10 ⁴
	CNR 92 J6	<i>E. cloacae</i>	0.25	1	0.12	■■■ Case	> 1 x 10 ⁴	> 1 x 10 ⁴
	2 F10	<i>E. cloacae</i>	0.12	1	0.12	■■■ Case	1 x 10 ³	> 1 x 10 ⁴
	2 G1	<i>E. cloacae</i>	0.12	1	0.12	■■■ Case	1 x 10 ²	> 1 x 10 ⁴
	2 G2	<i>E. cloacae</i>	0.25	4	0.25	■■■ Case	1 x 10 ²	2 x 10 ³
	2 G3	<i>E. cloacae</i>	4	1.5	0.75	■■■ Case	5 x 10 ³	> 1 x 10 ⁴
	2 G4	<i>E. cloacae</i>	0.19	1.5	0.12	■■■ Case	1 x 10 ²	> 1 x 10 ⁴
	2 G5	<i>E. cloacae</i>	0.5	4	0.75	■■■ Case	1 x 10 ²	> 1 x 10 ⁴
	2 G6	<i>E. cloacae</i>	1.5	0.75	0.25	■■■ Case	1 x 10 ²	> 1 x 10 ⁴
	2 G7	<i>E. cloacae</i>	1.5	2	0.75	■■■ Case	1 x 10 ²	> 1 x 10 ⁴
	2 G8	<i>E. cloacae</i>	0.5	1.5	0.38	■■■ Case	1 x 10 ²	> 1 x 10 ⁴
	2 G9	<i>E. cloacae</i>	2	4	1.5	■■■ Case	1 x 10 ²	1 x 10 ³
	2 G10	<i>E. cloacae</i>	8	>32	4	■■■ Case	1 x 10 ²	4 x 10 ²
	2 H1	<i>E. cloacae</i>	0.5	4	0.75	■■■ Case	1 x 10 ²	1 x 10 ²
	2 H2	<i>E. cloacae</i>	0.19	3	0.38	■■■ Case	1 x 10 ²	1 x 10 ²
	2 H3	<i>E. cloacae</i>	1.5	3	0.75	■■■ Case	1 x 10 ²	> 1 x 10 ⁴
	2 H4	<i>E. aerogenes</i>	1	4	0.75	■■■ Case	1 x 10 ²	> 1 x 10 ⁴
	2 H5	<i>M. morganii</i>	1.5	0.02	0.12	■■■ Case	> 1 x 10 ⁴	> 1 x 10 ⁴

CNR 87 J10	<i>H. alvei</i>	0.25	1	0.09	☒☒☒ Case	> 1 x 10 ⁴	> 1 x 10 ⁴	
CNR 88 A1	<i>H. alvei</i>	0.38	6	0.75	☒☒☒ Case	> 1 x 10 ⁴	1 x 10 ²	
CNR 87 F7	<i>S. marcescens</i>	0.75	0.75	0.19	☒☒☒ Case	1 x 10 ²	1 x 10 ²	
CNR 87 B7	<i>E. cloacae</i>	0.25	0.75	0.04	☒☒☒ Case	1 x 10 ²	> 1 x 10 ⁴	
ESBL + impermeability	2 H6	<i>E. coli</i>	2	4	1	CTX-M-15	1 x 10 ²	1 x 10 ²
	2 H7	<i>K. pneumoniae</i>	1	>32	4	CTX-M-15 + SHV-1	1 x 10 ²	1 x 10 ²
	2 H8	<i>K. pneumoniae</i>	1.5	>32	4	CTX-M-15 + TEM-1 + SHV-1	1 x 10 ²	1 x 10 ²
	2 H9	<i>K. pneumoniae</i>	0.25	1	1	CTX-M-15 + TEM-1 + SHV-1	1 x 10 ²	> 1 x 10 ⁴
	2 H10	<i>K. pneumoniae</i>	1.5	>32	6	CTX-M-15 + SHV-11	1 x 10 ²	1 x 10 ²
	2 I1	<i>K. pneumoniae</i>	8	>32	4	CTX-M-15 + SHV-28 - TEM-1	1 x 10 ²	1 x 10 ²
	2 I2	<i>K. pneumoniae</i>	1	4	1	TEM-1 + SHV-28	1 x 10 ²	1 x 10 ²
	2 I3	<i>K. pneumoniae</i>	3	>32	6	CTX-M-15 + TEM-1 + SHV-11	1 x 10 ²	1 x 10 ²
	2 I4	<i>K. pneumoniae</i>	0.25	1	1	CTX-M-15 + TEM-1 + SHV-11	1 x 10 ²	1 x 10 ²
	2 I5	<i>K. pneumoniae</i>	6	> 32	> 32	CTX-M-15 + TEM-1 + SHV-11	1 x 10 ²	1 x 10 ²
	2 I6	<i>K. pneumoniae</i>	0.75	> 32	3	CTX-M-15 + TEM-1 + SHV-12	1 x 10 ²	1 x 10 ²
	2 I7	<i>K. pneumoniae</i>	1	24	0.5	CTX-M-15 + TEM-1 + SHV-11	1 x 10 ²	1 x 10 ²
	2 I8	<i>K. pneumoniae</i>	2	4	1	CTX-M-15 + TEM-1 + SHV-1 + OXA-1	1 x 10 ²	> 1 x 10 ⁴
Hyper K1 + impermeability	CNR 151 J9	<i>K. oxytoca</i>	2	3	0.75	☒☒☒ K1 penicillinase	1 x 10 ²	1 x 10 ²
	CNR 92 J7	<i>K. oxytoca</i>	1	2	0.5	☒☒☒ K1 penicillinase	1 x 10 ²	> 1 x 10 ⁴
ESBL + Case + impermeability	2 I9	<i>E. cloacae</i>	1.5	6	1	☒☒☒ Case + CTX-M-15	1 x 10 ²	> 1 x 10 ⁴
	2 I10	<i>E. cloacae</i>	2	8	1	☒☒☒ Case + CTX-M-15	1 x 10 ²	> 1 x 10 ⁴
	2 J1	<i>E. cloacae</i>	3	12	2	☒☒☒ Case + CTX-M-15	1 x 10 ²	> 1 x 10 ⁴
	2 J2	<i>C. freundii</i>	1	8	1	☒☒☒ Case + TEM-3	1 x 10 ²	1 x 10 ²
Extended spectrum oxacillinases	2 J3	<i>K. pneumoniae</i>	0.5	0.38	0.12	OXA-163	> 1 x 10 ⁴	> 1 x 10 ⁴
	2 J4	<i>E. cloacae</i>	0.5	2	0.19	OXA-163	2 x 10 ²	> 1 x 10 ⁴
	2 J5	<i>S. marcescens</i>	0.5	0.75	0.19	OXA-405	> 1 x 10 ⁴	> 1 x 10 ⁴

CARBAPENEMASE PRODUCERS

KPC-type	1 F1	<i>E. coli</i>	1	>32	3	KPC-2	1 x 10 ²	1 x 10 ²
	1 F2	<i>E. coli</i>	0.5	0.5	0.5	KPC-2	1 x 10 ²	1 x 10 ²
	1 F3	<i>E. coli</i>	2	1.5	1	KPC-2 + TEM-1 + OXA-9	1 x 10 ²	1 x 10 ²
	1 F4	<i>E. coli</i>	4	4	2	KPC-2 + CTX-M-9 + TEM-1	1 x 10 ²	4 x 10 ²
	1 F5	<i>K. pneumoniae</i>	16	24	32	KPC-2 + SHV-11 + TEM-1 + CTX-M-2	1 x 10 ²	1 x 10 ²
	1 F6	<i>K. pneumoniae</i>	>32	>32	>32	KPC-2 + SHV-11 + TEM-1 + CTX-M-2 + OXA-9	1 x 10 ²	1 x 10 ²
	1 F7	<i>K. pneumoniae</i>	16	>32	>32	KPC-2 + SHV-11 + CTX-M-15	1 x 10 ²	1 x 10 ²
	1 F8	<i>K. pneumoniae</i>	4	4	32	KPC-2 + TEM-1 + SHV-1 + CTXM-15	1 x 10 ²	1 x 10 ²
	1 F9	<i>K. pneumoniae</i>	4	24	2	KPC-2 + SHV-11 + TEM-1 + SHV-12 + OXA-9	1 x 10 ²	1 x 10 ²
	1 F10	<i>K. pneumoniae</i>	>32	>32	>32	KPC-2 + SHV-11	1 x 10 ²	1 x 10 ²
	1 G1	<i>K. pneumoniae</i>	4	6	8	KPC-2 + SHV-11 + TEM-1	1 x 10 ²	1 x 10 ²
	1 G2	<i>K. pneumoniae</i>	4	>32	8	KPC-3 + TEM-1 + SHV-1 + CTX-M-15 + OXA-9	1 x 10 ²	1 x 10 ²
	1 G3	<i>K. pneumoniae</i>	8	>32	8	KPC-3 + SHV-11 + OXA-9 + TEM-1	1 x 10 ²	1 x 10 ²
	1 G4	<i>K. ozoenae</i>	>32	>32	2	KPC-3 + OXA-9 + TEM-1	> 1 x 10 ⁴	2 x 10 ²

1	G5	<i>E. cloacae</i>	1	1.5	0.75	KPC-2	1×10^2	1×10^2
1	G6	<i>E. cloacae</i>	24	>32	16	KPC-2 + TEM-1	1×10^2	1×10^2
1	G7	<i>E. cloacae</i>	4	6	2	KPC-2 + TEM-1 + OXA-1	1×10^2	1×10^2
1	G8	<i>E. cloacae</i>	2	4	1.5	KPC-2 + TEM-1 + SHV-11	1×10^2	1×10^2
1	G9	<i>E. cloacae</i>	2	2	1	KPC-2 + TEM-3	1×10^2	1×10^2
1	G10	<i>C. freundii</i>	8	1.5	3	KPC-2 + TEM-1	$> 1 \times 10^4$	1×10^2
1	I1	<i>S. marcevens</i>	>32	>32	>32	KPC-2 + TEM-1 + SHV-12	1×10^2	1×10^2
1	I2	<i>S. marcevens</i>	>32	>32	>32	KPC-2 + TEM-1	1×10^2	1×10^2
IMI-type	I3	<i>E. cloacae</i>	>32	8	4	IMI-1	$> 1 \times 10^4$	1×10^2
	I4	<i>E. asburiae</i>	>32	>32	>32	IMI-2	$> 1 \times 10^4$	1×10^2
	I5	<i>E. asburiae</i>	>32	8	2	IMI-2	$> 1 \times 10^4$	3×10^2
NMC-A	I6	<i>E. cloacae</i>	>32	1,5	0,75	NmcA	$> 1 \times 10^4$	3×10^3
SME-type	I7	<i>S. marcescens</i>	>32	8	8	Sme-1	$> 1 \times 10^4$	$> 1 \times 10^4$
	I8	<i>S. marcescens</i>	>32	8	8	Sme-2	$> 1 \times 10^4$	1×10^2
GES-type	I9	<i>E. cloacae</i>	>32	>32	>32	GES-5	1×10^2	1×10^2
FRI-1	I10	<i>E. cloacae</i>	>32	>32	16	FRI-1	1×10^2	1×10^2
NDM-type	A1	<i>E. coli</i>	1	3	1	NDM-1 + OXA-1 + OXA-10 + CMY-16 + TEM-1	1×10^2	1×10^2
	A2	<i>E. coli</i>	3	3	2	NDM-1 + OXA-1 + TEM-1	1×10^2	1×10^2
	A3	<i>E. coli</i>	6	32	16	NDM-1 + CTX-M-15 + TEM-1	1×10^2	1×10^2
	A4	<i>E. coli</i>	4	>32	8	NDM-1 + OXA-1 + OXA-2 + CTX-M-15 + TEM-1	1×10^2	1×10^2
	A5	<i>E. coli</i>	16	>32	16	NDM-1 + CTX-M-15 + TEM-1	1×10^2	1×10^2
	A6	<i>E. coli</i>	>32	>32	>32	NDM-4 + CTX-M-15 + OXA-1	1×10^2	1×10^2
	A7	<i>E. coli</i>	>32	>32	>32	NDM-4 + CTX-M-15 + CMY-6	1×10^2	1×10^2
	A8	<i>E. coli</i>	>32	>32	>32	NDM-5 + TEM-1 + CTX-M-15	1×10^2	1×10^2
	A9	<i>E. coli</i>	6	32	8	NDM-6 + CTX-M-15 + OXA-1	1×10^2	1×10^2
	A10	<i>E. coli</i>	4	16	3	NDM-7	1×10^2	1×10^2
	B1	<i>K. pneumoniae</i>	2	8	3	NDM-1 + CTX-M-15 + SHV-11 + OXA-1	1×10^2	1×10^2
	B2	<i>K. pneumoniae</i>	>32	>32	>32	NDM-1 + CTX-M-15 + CMY-4 + OXA-1	1×10^2	1×10^2
	B3	<i>K. pneumoniae</i>	>32	>32	>32	NDM-1 + CTX-M-15 + OXA-1 + OXA-9 + TEM-1 + SHV-28 + SHV-11	1×10^2	1×10^2
	B4	<i>K. pneumoniae</i>	1,5	6	2	NDM-1 + OXA-1 + SHV-11	1×10^2	1×10^2
	B5	<i>K. pneumoniae</i>	1	8	4	NDM-1 + OXA-1 + CTX-M-15 + TEM-1 + SHV-28 + OXA-9 + CMY-6	1×10^2	1×10^2
	B6	<i>K. pneumoniae</i>	1,5	8	1,5	NDM-1 + TEM-1 + CTX-M-15 + SHV-12 + OXA-9	1×10^2	1×10^2
	B7	<i>K. pneumoniae</i>	4	8	16	NDM-1 + TEM-1 + CTX-M-15 + SHV-12 + OXA-9	1×10^2	1×10^2
	B8	<i>K. pneumoniae</i>	2	>32	4	NDM-1 + TEM-1 + CTX-M-15 + SHV-11 + OXA-1	1×10^2	1×10^2
	B9	<i>P. stuartii</i>	12	0,38	1,5	NDM-1 + OXA-1 + CMY-6 + TEM-1	1×10^2	$> 1 \times 10^4$
	B10	<i>P. rettgeri</i>	3	0,5	1,5	NDM-1 + CTX-M-15	1×10^2	1×10^2
	C1	<i>Salmonella enterica</i>	4	6	3	NDM-1 + CTX-M-15 + TEM-1 + OXA-1 + OXA-9 + OXA-10	5×10^2	2×10^2
VIM-type	C2	<i>E. coli</i>	1,5	0,38	0,5	VIM-1 + CTX-M-3	1×10^2	$> 1 \times 10^4$
	C3	<i>E. coli</i>	3	1,5	1	VIM-1 + CMY-13	2×10^3	6×10^2
	C4	<i>E. coli</i>	8	4	3	VIM-4	1×10^2	1×10^2
	C5	<i>K. pneumoniae</i>	>32	>32	>32	VIM-1 + SHV-5	1×10^2	1×10^2

1	C6	<i>K. pneumoniae</i>	>32	>32	>32	VIM-1		1 x 10 ²	1 x 10 ²
1	C7	<i>K. pneumoniae</i>	>32	>32	>32	VIM-1 + SHV-12		1 x 10 ²	1 x 10 ²
1	C8	<i>K. pneumoniae</i>	>32	>32	>32	VIM-1		1 x 10 ²	1 x 10 ²
1	C9	<i>K. pneumoniae</i>	4	2	2	VIM-1 + SHV-5		1 x 10 ²	1 x 10 ²
1	C10	<i>K. pneumoniae</i>	>32	>32	>32	VIM-1 + TEM-1 + SHV-5		1 x 10 ²	1 x 10 ²
1	D1	<i>K. pneumoniae</i>	>32	>32	>32	VIM-1 + SHV-5		1 x 10 ²	1 x 10 ²
1	D2	<i>K. pneumoniae</i>	1	0.5	1	VIM-1 + CTX-M-3		1 x 10 ²	1 x 10 ²
1	D3	<i>K. pneumoniae</i>	0.5	4	0.38	VIM-1 + SHV-5		1 x 10 ²	1 x 10 ²
1	D4	<i>K. pneumoniae</i>	8	16	4	VIM-19 + CTX-M-3 + TEM-1 + SHV-1		1 x 10 ²	1 x 10 ²
1	D5	<i>E. cloacae</i>	1	0.38	0.5	VIM-1 + SHV-70		1 x 10 ²	3 x 10 ²
1	D6	<i>E. cloacae</i>	3	2	1	VIM-4 + CTX-M-15 + TEM-1 + SHV-31		1 x 10 ²	1 x 10 ²
1	D7	<i>C. freundii</i>	2	2	0.75	VIM-2 + TEM-1 +		1 x 10 ²	1 x 10 ²
1	D8	<i>C. freundii</i>	1.5	4	0.5	VIM-2 + TEM-1 + OXA-9 + OXA-10		1 x 10 ²	1 x 10 ²
	D9	<i>E. coli</i>	0.5	3	0.5	IMP-1		1 x 10 ²	1 x 10 ²
	D10	<i>E. coli</i>	6	8	3	IMP-8 + SHV -12		1 x 10 ²	1 x 10 ²
	E1	<i>K. pneumoniae</i>	1.5	3	1	IMP-1		1 x 10 ²	1 x 10 ²
	E2	<i>K. pneumoniae</i>	8	3	2	IMP-1 + TEM-15		1 x 10 ²	1 x 10 ²
	E3	<i>K. pneumoniae</i>	1.5	4	2	IMP-1 + TEM-15 + CTX-M-15		1 x 10 ²	1 x 10 ²
IMP-type	E4	<i>K. pneumoniae</i>	1	2	8	IMP-1 + SHV-5		1 x 10 ²	1 x 10 ²
	E5	<i>K. pneumoniae</i>	1	1	0.5	IMP-8		1 x 10 ²	1 x 10 ²
	E6	<i>K. pneumoniae</i>	0.5	0.5	0.5	IMP-8 + SHV -12		1 x 10 ²	1 x 10 ²
	E7	<i>E. cloacae</i>	1.5	1	1	IMP-8		1 x 10 ²	1 x 10 ²
	E8	<i>E. cloacae</i>	0.75	0.5	0.5	IMP-8 + SHV-12		1 x 10 ²	1 x 10 ²
	E9	<i>S. marscevens</i>	8	>32	2	IMP-11		1 x 10 ²	1 x 10 ²
GIM-type	E10	<i>E. cloacae</i>	2	>32	6	GIM-1		1 x 10 ²	1 x 10 ²
	A1	<i>E. coli</i>	3	16	1	OXA-48 + CTX-M-15		1 x 10 ²	1 x 10 ²
	A2	<i>E. coli</i>	0.5	0.75	0.12	OXA-48 + CTX-M-15		1 x 10 ²	1 x 10 ²
	A3	<i>E. coli</i>	0.38	1.5	0.19	OXA-48 + CTX-M-15		1 x 10 ²	1 x 10 ²
	A4	<i>E. coli</i>	0.25	0.5	0.19	OXA-48 + CTX-M-24 + TEM-1		1 x 10 ²	1 x 10 ²
	A5	<i>E. coli</i>	0.75	1	0.19	OXA-48 + CTX-M-24 + TEM-1		1 x 10 ²	1 x 10 ²
	A6	<i>E. coli</i>	0.5	1	0.25	OXA-48 + CTX-M-15		1 x 10 ²	2 x 10 ²
	A7	<i>K. pneumoniae</i>	0.5	2	0.5	OXA-48		2 x 10 ³	1 x 10 ²
OXA-48	A8	<i>K. pneumoniae</i>	0.38	1	0.5	OXA-48 + TEM-1		> 1 x 10 ⁴	1 x 10 ²
	A9	<i>K. pneumoniae</i>	2	3	2	OXA-48 + CTX-M-15		1 x 10 ²	1 x 10 ²
	A10	<i>K. pneumoniae</i>	1	4	1	OXA-48		> 1 x 10 ⁴	1 x 10 ²
	B1	<i>K. pneumoniae</i>	1	4	1	OXA-48		> 1 x 10 ⁴	1 x 10 ²
	B2	<i>K. pneumoniae</i>	>32	>32	>32	OXA-48		6 x 10 ³	1 x 10 ²
	B3	<i>K. pneumoniae</i>	0.5	0.75	0.25	OXA-48 + SHV-11		> 1 x 10 ⁴	1 x 10 ²
	B4	<i>K. pneumoniae</i>	0.5	>32	1,5	OXA-48 + CTX-M-15 + TEM -1		1 x 10 ²	1 x 10 ²
	B5	<i>K. pneumoniae</i>	0.75	16	1	OXA-48 + CTX-M-15 + TEM-1		1 x 10 ²	1 x 10 ²
	B6	<i>E. cloacae</i>	0.5	2	0.5	OXA-48 + TEM-1 + CTX-M-15 + OXA-1		1 x 10 ²	1 x 10 ²

	2	B7	<i>E. cloacae</i>	1	16	1.5	OXA-48 + TEM-1 + CTX-M-15 + OXA-1	1×10^2	1×10^2
	2	B8	<i>E. cloacae</i>	2	8	1	OXA-48 + SHV-5	1×10^2	1×10^2
	2	B9	<i>C. koseri</i>	0.38	2	0.38	OXA-48	1×10^3	1×10^2
	2	B10	<i>C. koseri</i>	0.75	2	0.38	OXA-48 + TEM-1	1×10^2	1×10^2
	2	C1	<i>C. freundii</i>	0.75	1,5	0.38	OXA-48 + SHV-12 + TEM-1	1×10^2	1×10^2
OXA-162	2	C2	<i>K. pneumoniae</i>	4	8	1	OXA-162 + TEM-1 + SHV-11	1×10^2	1×10^2
		CNR 67 F9	<i>K. pneumoniae</i>	0.38	8	0.5	OXA-162	$> 1 \times 10^4$	1×10^2
OXA-181	2	C3	<i>E. coli</i>	0.38	1.5	0.19	OXA-181	1×10^3	1×10^2
	2	C4	<i>E. coli</i>	0.5	1.5	0.25	OXA-181	$> 1 \times 10^4$	1×10^4
		CNR 51 E10	<i>E. coli</i>	0.38	0.38	0.12	OXA-181	1×10^2	1×10^4
		CNR 59 F5	<i>E. coli</i>	0.38	1.5	0.12	OXA-181	1×10^2	1×10^2
		CNR 61 C6	<i>E. coli</i>	0.25	1	0.12	OXA-181	1×10^2	1×10^2
		CNR 61 D1	<i>E. coli</i>	0.25	1.5	0.12	OXA-181	1×10^2	1×10^2
		CNR 64 C4	<i>E. coli</i>	>32	>32	12	OXA-181	1×10^2	1×10^2
	2	C9	<i>K. pneumoniae</i>	0.5	2	0.5	OXA-181 + SHV-11 + CTXM-15 + OXA-1	1×10^2	1×10^2
		CNR 58 J8	<i>K. pneumoniae</i>	0.38	1.5	0.38	OXA-181	1×10^2	1×10^2
OXA-204	2	D2	<i>K. pneumoniae</i>	0.5	2	0.5	OXA-204 + CMY-4	1×10^2	1×10^2
	2	D3	<i>E. coli</i>	0.38	0.19	0.094	OXA-204 + CMY-2 + CTX-M-15 + OXA-1	1×10^2	1×10^2
	2	D4	<i>E. coli</i>	0,5	2	0.25	OXA-204 + CMY-4+ CTX-M-15 + OXA-1	1×10^2	1×10^4
	2	D5	<i>E. coli</i>	0,5	2	0.38	OXA-204 + CMY-4 + CTX-M-15	1×10^2	1×10^3
	2	D6	<i>K. pneumoniae</i>	0,5	16	0.75	OXA-204 + SHV-28 + TEM-1 + CTX-M-15	1×10^2	1×10^2
OXA-232	2	D7	<i>E. coli</i>	>32	>32	>32	OXA-232 + CTX-M-15 + OXA-1	1×10^2	1×10^2
	2	D8	<i>K. pneumoniae</i>	3	>32	12	OXA-232 +SHV-1 + TEM-1 + CTX-M-15 + OXA-1	1×10^2	1×10^2
		CNR 68 B2	<i>E. coli</i>				OXA-232	$> 1 \times 10^4$	1×10^2
OXA-244	2	D9	<i>E. coli</i>	0.5	2	0.5	OXA-244 + TEM-1 + CMY-2	$> 1 \times 10^4$	1×10^2
	2	D10	<i>E. coli</i>	0.5	1.5	0.5	OXA-244 + TEM-1 + CMY-2	1×10^2	1×10^2
Multiple carbapenemases	2	J6	<i>C. freundii</i>	3	2	0.5	OXA-372 + CMY-135 + OXA-10 + MOX-9	1×10^2	1×10^2
	2	C5	<i>K. pneumoniae</i>	3	>32	4	OXA-181 + SHV-11 + TEM-1 + CTX-M-15 + NDM-1 + OXA-1	1×10^2	1×10^2
	2	C6	<i>K. pneumoniae</i>	>32	>32	>32	OXA-181 + SHV-27 + CTX-M-15 + TEM-1 + NDM-1 + OXA-1	1×10^2	1×10^2
	2	C7	<i>K. pneumoniae</i>	>32	>32	>32	OXA-181 + SHV-11 + CTX-M-15 + NDM-1 + OXA-1	1×10^2	1×10^2
	2	C8	<i>K. pneumoniae</i>	16	>32	>32	OXA-181 + SHV-11 + TEM-1 + CTX-M-15 + NDM-1 + OXA-9	1×10^2	1×10^2
	2	C10	<i>K. pneumoniae</i>	0.5	2	0.5	OXA-181 + NDM-1 + SHV-2 + CTX-M-15 + OXA-1	1×10^2	1×10^2
	2	D1	<i>C.freundii</i>	>32	>32	16	OXA-181 + NDM-1 + OXA-1 + OXA-9 + OXA-10 + CTX-M-15 + TEM-1	1×10^2	1×10^2
		CNR 60 A6	<i>E. coli</i>	16	>32	16	NDM-1 + OXA-48	1×10^2	1×10^2
		CNR 63 C5	<i>E. coli</i>	2	>32	3	NDM-1 + OXA-48	1×10^2	1×10^2
		CNR 67 D2	<i>E. coli</i>	1.5	>32	3	NDM-1 + OXA-48	1×10^2	1×10^2
		CNR 22 F9	<i>K. pneumoniae</i>	4	>32	8	NDM-1 + OXA-232	1×10^2	1×10^2
		CNR 46 A8	<i>E. coli</i>	2	>32	6	NDM-1 + OXA-232	1×10^2	1×10^2
		CNR 28 10	<i>K. pneumoniae</i>	>32	>32	>32	NDM-5 + OXA-232	1×10^2	1×10^2
		CNR 51 A9	<i>E. coli</i>	0.75	1	0.38	NDM-1 + VIM-2	1×10^2	1×10^2
		CNR 45 J4	<i>E. cloacae</i>	1	3	0.75	VIM-4 + OXA-48	1×10^2	1×10^2

^a Abbreviations : IPM, imipenem; ETP, ertapenem, MEM, meropenem, ~~AAA~~ Case, overexpressed cephalosporinase

^b One ml of stools contains 100 mg of stool.

^c Underlined CFU counts are considered as negative results (cut-off values set at $\geq 1 \times 10^4$ CFU/ ml).