

CHROMagar™ ESB^L

Instructions For Use

For Research Use Only (RUO).

Not for use in diagnostic procedures.

Chromogenic medium for overnight detection of gram-negative bacteria producing extended spectrum beta-lactamase.

REFERENCES



Pack Size

Ordering References

Base (RT)

Supplement (ES)

5000 mL

250 Tests
of 20 mL

=

ESRT2

=

RT412

Weight: 165 g

+

ES372

Weight: 2,85 g

25 L

1250 Tests
of 20 mL

=

ESRT3-25

=

RT413-25

Weight: 825 g

+

ES373-25

Weight: 14,25 g

INTENDED USE

CHROMagar™ ESB^L is a selective and differential chromogenic culture medium, intended for use in the qualitative direct detection of gastrointestinal colonization with Extended-spectrum-beta-lactamase-resistant Enterobacteria (ESBL) to aid in the prevention and control of ESB^L in healthcare settings. The test is performed with rectal swab and stools from patients to screen for ESB^L colonization. Results can be interpreted after 18-24 h of aerobic incubation at 35-37 °C.

The medium can also be used as an early warning indicator for diagnostic tests of infections to signal the possible presence of multi drug-resistant bacteria. This use does not replace the institution's protocols. CHROMagar™ ESB^L is not intended to diagnose infection caused by ESB^L producers nor to guide nor monitor treatment for infections. A lack of growth or the absence of colonies on CHROMagar™ ESB^L does not preclude the presence of ESB^L-producing bacteria. Further identification, susceptibility testing, and epidemiological typing is needed on suspect colonies.

COMPOSITION

The product is composed of a powder base (CHROMagar™ Orientation) and 1 supplement (CHROMagar™ ESB^L supplement).

Product	=	Base (RT)	+	Supplement (ES)
Total g/L		33.0 g/L		0.57 g/L
Composition g/L		Agar 15.0 Peptone and yeast extract 17.0 Chromogenic mix 1.0		Selective mix 0.57
Aspect		Powder Form		Powder Form
STORAGE		15-30 °C		2/8 °C

Need some
Technical Documents?

Available
for download on
www.CHROMagar.com

- Certificate of Analysis (CoA) --> One per Lot
- Material Safety Data Sheet (MSDS)

FINAL MEDIA pH

7.0 +/- 0.2

PREPARATION (Calculation for 1 L)

Step 1

Preparation of the base
CHROMagar™
Orientation

- Disperse slowly 33 g of powder base in 1 L of purified water.
- Stir until agar is well thickened.
- Heat and bring to boiling (100 °C) while swirling or stirring regularly.

Advice 1: For enhanced growth, add 0.5 g/L of Tween 80 to the previous preparation mix.

Advice 2: For the 100 °C heating step, mixture may also be brought to a boil in a microwave oven: after initial boiling, remove from oven, stir gently, then return to oven for short repeated bursts of heating until complete fusion of the agar grains has taken place (large bubbles replacing foam).

Step 2

Autoclave

- AUTOCLAVE at 121 °C during 15 min.
- Cool in a water bath to 45/50 °C, swirling or stirring gently.

Step 3

Preparation
of the
CHROMagar™ ESB^L
supplement

- Weight 570 mg of the required supplement powder.
- Add 10 mL of purified sterile water to this powder to make a supplement solution.

Warning 1: This step may require several minutes of stirring to obtain a good and homogenous suspension: **opaque yellowish appearance**.

Warning 2: Reconstituted supplement solution must be used the same day.

Warning 3: Do not store and re-use a supplement solution.

Final
Media

HELPING CALCULATION

1 L

Rehydrate 570 mg
into 10 mL of purified water

5 L

Rehydrate 2,85 g
into 50 mL of purified water

25 L

Rehydrate 14,25 g
into 250 mL of purified water

Step 4

Integrate the
supplement to the
melted base

- Vortex this supplement to homogenize and add this supplement solution to melted CHROMagar™ Orientation cooled at 45/50 °C.
- Stir to make CHROMagar™ ESB^L.

Step 5

Pouring

- Pour into sterile Petri dishes.
- Let it solidify and dry.

Storage

- Store in the dark before use.
- Prepared media plates can be kept for one day at room temperature.
- Plates can be stored for up to 1 month under refrigeration (2/8 °C) if properly prepared and protected from light and dehydration.

SPECIMEN COLLECTION AND HANDLING

CHROMagar™ ESB can be used with the following specimens:
Rectal swabs and stools

Sampling and transport equipment must be used in accordance with the recommendations of their suppliers for the conservation of ESB strains.

MATERIAL REQUIRED BUT NOT PROVIDED

Standard microbiological laboratory material for culture media preparation, control, streaking, incubation and waste disposal.

INOCULATION

Related samples are inoculated by direct streaking on the plate.


- If the agar plate has been refrigerated, allow to warm to room temperature before inoculation.
- Streak sample onto plate.
- Incubate in aerobic conditions at 35-37 °C for 18-24 hours.

INTERPRETATION

Qualitative reading and interpretation of the petri dishes

Microorganism	Typical colony appearance
ESBL <i>E. coli</i>	→ dark pink to reddish
ESBL KEC (<i>Klebsiella</i> , <i>Enterobacter</i> , <i>Citrobacter</i>)	→ metallic blue (+/- reddish halo)
ESBL <i>Proteus</i>	→ brown halo
ESBL <i>Acinetobacter</i>	→ cream
ESBL <i>Pseudomonas</i>	→ translucent, (+/- natural pigmentation cream to green)
<i>Stenotrophomonas</i>	→ colourless
Gram (+) strains	→ inhibited
Non Resistant Other Gram (-) strains	→ inhibited
Yeasts	→ mostly inhibited

Typical colony appearance



The pictures shown are not contractual.

PERFORMANCE

	Analytical data *	Clinical data **	
	CHROMagar™ ESBL	CHROMagar™ ESBL	Competitor's medium
Sensitivity	100 %	98.3 %	97.5 %
Specificity	97 %	72.3 %	72.9 %

* Data obtained after a 18-24 h incubation at 37 °C in aerobic conditions in the study « Rapid detection of ESB-producing Enterobacteria on chromogenic medium: Colorex Orientation/ESBL ». Laudat et al. Poster SFM 2010.

** Data obtained by testing 2 337 rectal screening samples, on plates incubated at 37 °C for 18 h, in the study « Comparison of five media for detection of Extended-Spectrum beta-Lactamase by use of the wrap instrument for automated specimen processing ». Grohs et al., 2013. *J.Clin. Microbiol.*

LIMITATIONS AND COMPLEMENTARY TESTS

- Some *Pseudomonas* spp and *Acinetobacter* spp, widely-known to be frequently Multi Drug Resistant bacteria, could grow on the medium with typical colony aspects as it is observed on CHROMagar™ Orientation.
- Final identification may require additional testing such as biochemical tests or mass spectrometry (e.g MALDI-TOF) which can be done directly from the suspicious colonies observed on the medium.
- Most AmpC-producing bacteria are inhibited but some may show some growth.
- The antibiotic resistant mechanism must be confirmed by recommended tests (e.g CLSI, EUCAST).

QUALITY CONTROL

Please perform Quality Control according to the use of the medium and the local QC regulations and norms.

Good preparation of the medium can be tested, isolating the following ATCC strains:

Microorganism	Typical colony appearance
ESBL <i>E. coli</i> CIP 103982	→ reddish, small colonies
ESBL <i>K. pneumoniae</i> ATCC® 700603	→ metallic blue
<i>E. faecalis</i> ATCC® 29212	→ inhibited
<i>P. aeruginosa</i> ATCC® 10145	→ inhibited
<i>E. coli</i> ATCC® 25922	→ inhibited
<i>C. albicans</i> ATCC® 60193	→ inhibited
<i>S. aureus</i> ATCC® 25923	→ inhibited

WARNINGS AND PRECAUTIONS

- For Research Use Only (RUO). Not for use in diagnostic procedures.
- This laboratory product should be used only by trained personnel (healthcare professional, etc). Wear appropriate protective clothing, gloves and eye/face protection and handle appropriately with procedures and good laboratory practices.
- Use of the medium may be difficult for people who have problems recognising colours.
- Culture media should not be used as manufacturing material or components.
- Do not ingest or inhale the product.
- Do not use the product after the expiry date.
- Do not use the product if it shows any evidence of contamination or any sign of deterioration (compacted powder, color change, ...).
- Do not use the product if the packaging is damaged.
- Any change or modification in the production procedure may affect the results.
- Any change or modification of the required storage temperature may affect the performance of the product.
- Unappropriate storage may affect the shelf life of the product.

- Recap the bottles/vials tightly after each preparation and keep them in a low humidity environment, protected from moisture and light.
- Do not use the culture medium poured into a Petri dish after a first use.
- After opening the bottles and with an appropriate conservation, open bottles can be used under the same conditions until each product's expiry date.
- Reading and interpretation should be performed using isolated colonies.
- Some precipitate may be observed in the agar but these do not affect the performance of the product.
- Interpretation of the test results should be made taking into consideration colonial and microscopic morphology and the results of any other tests performed.
- Laboratory, chemical or biohazardous wastes must be handled and discarded in accordance with all local and national regulations.
- For hazard and precaution recommendations related to some chemical components in this medium, please refer to the pictogram(s) mentioned on the labels. The Material Safety Data Sheet (MSDS) is available on www.chromagar.com
- Any incident or complaint related to the environment must be declared to the manufacturer at the following email address: chromagar@chromagar.com
- Any serious incident occurring in connection with the environment must be declared to the competent authorities and to the manufacturer at the following email address: chromagar@chromagar.com

DISPOSAL OF WASTE









After use, all plates and any other contaminated materials must be sterilized or disposed of by appropriate internal procedures and in accordance with local legislations. Plates can be destroyed by autoclaving at 121 °C for at least 20 minutes.

LITERATURE REFERENCES

Please refer to our website page «Publications» for scientific publications about this particular product.

Web link: <http://www.chromagar.com/publication.php>

IFU/LABEL INDEX

-  Catalogue reference
-  Consult instructions for use
-  Quantity of powder sufficient for X liters of media
-  Expiry date
-  Required storage temperature
-  Store away from humidity
-  Protect from light
-  Manufacturer

NT-EXT-50 USA V7.0 / 27-May-22

CHROMagar™ and Rambach™ are trademarks created by Dr A. Rambach
ATCC® is a registered trademark of the American Type Culture Collection